

NeuroMetrix DPNCheck


DPNCheck 2.0 Device Training



Commercial stage neurotechnology company

- Based in Woburn, MA
- Trade on Nasdaq (NURO)
- Over 5M patients served
- Trusted by largest VBC providers and MA payors
- Three commercial products
- Extensive IP portfolio
- Fully integrated operations

DPNCheck®



ADVANCE®



Quell™



Our mission is to improve patient outcomes and population health by detecting, quantifying, and helping providers to reduce the impact of neurological disorders

Polyneuropathy Background



Polyneuropathy is common and leads to substantial morbidity and mortality

- Clinical testing has low sensitivity for early-stage disease
- Laboratory testing is expensive, only appropriate for confirmation
- Unmet need for accurate, widely available, screening test for polyneuropathy

Prevalence*

10% in overall population

Up to 30%+ in Medicare Population



Multiple Causes



High Prevalence



Serious Complications

*Mold et al. 2004. Hanewinkel et al. 2016. Singer et al. 2012. Dyck et al. 1993.


Etiology of Peripheral Neuropathies

- **Metabolic**
 - Diabetes (30% of patients)
 - Metabolic syndrome (IGT, hypertension, dyslipidemia, obesity)
 - B12 deficiency
 - Thyroid disease
- **Chronic Hypoxia**
 - Obstructive sleep apnea
 - COPD
- **Toxic**
 - Chemotherapy Induced Peripheral Neuropathy (CIPN)
 - Alcoholic neuropathy, uremic neuropathy
- **Inflammatory**
 - Rheumatoid arthritis
 - Chronic Inflammatory Demyelinating Polyneuropathy (CIDP)
 - Guillain-Barré syndrome (GBS)
- **Infectious**
 - HIV, Lyme disease
- **Hereditary**
 - Charcot-Marie-Tooth (CMT)



“Current evidence supports the association of the metabolic syndrome and its individual components with neuropathy.”


Some causes of neuropathy



Peripheral neuropathy has a high population prevalence

- 2x more common than PAD

Hicks et al. Sci Rep, 2021.: <https://doi.org/10.1038/s41598-021-98565-w>





NHANES: 1999–2004
National Health and
Nutrition Examination
Survey

ARIC: 2016–2017
Atherosclerosis Risk
in Communities Study

*Age, sex and race-adjusted prevalence of peripheral neuropathy stratified by diabetes status in US adults aged 40-69 and ≥ 70 Years (NHANES, 1999-2004) and ARIC participants aged ≥ 70 years (Visit 6, 2016-2017).

Peripheral neuropathy is an independent risk for falling and fall severity



OR, odds ratio.


References: Richardson and Hurvitz. J Gerontol, 1995. Cheng et al. J Clin Nurs, 2002. Erlandson et al. J Acquir Immune Defic Syndr, 2019. Riskowski et al. Journal of Foot and Ankle Research, 2012.

Elevated risk of falls and pain precede peripheral neuropathy Dx by several years

Callaghan et al. Neurology, 2015; <https://doi.org/10.1212/WNL.0000000000001714>

Figure 1 Comparison of the patient-oriented outcome trajectories between patients with neuropathy and propensity-matched controls

A. Have you fallen down in the last two years (or since previous wave)?




"We found that older adults with neuropathy have more falls and pain and lower self-rated health compared to carefully matched controls without neuropathy. These differences were present 3–5 years prior to a neuropathy diagnosis and persist for several years after diagnosis."

"This finding may be partly explained by a delay in diagnosis in this highly prevalent condition, and also highlights the fact that neuropathy often develops slowly over time. Patients typically report neuropathic symptoms to their physician years after their insidious onset."

Clinical screening tests have low sensitivity

- Monofilament only detects loss of protective sensation

Perkins et al. Diabetes Care, 2001.



*Corrected for false positive rate.

**Referenced publication utilized traditional NCS. DPNCheck sural nerve conduction demonstrated to have high agreement with traditional NCS.

Kural et al. 2018. Kamiya et al. 2021.
Scarr et al. 2018. Lee et al. 2014.

DPNCheck® Overview

DPNCheck is a standardized and automated sural nerve conduction test



- Performed in minutes by medical assistant
- Gold standard NCS technology
- Device + single-patient use biosensor
- High diagnostic accuracy
- Validated in 30+ peer-reviewed studies
- 2M patients tested over 10 years

American Diabetes Association Diabetes Care

Cancer Science JCA

THE LANCET
Diabetes & Endocrinology

nature
COMMUNICATIONS

MUSCLE
& NERVE

JAHA
Journal of the American Heart Association

frontiers
in Endocrinology

JDI Journal of Diabetes Investigation

Journal of Diabetes Research

PLOS ONE

CPIRPC

JDST
JOURNAL OF DIABETES
SCIENCE AND TECHNOLOGY

DIABETIC
Medicine

Respiratory Research

Journal of Diabetes Research

AMS
Archives of Medical Science

DIABETES
TECHNOLOGY & THERAPEUTICS

device display is simulated

Device Components



The device is powered by a 3V Panasonic CR123A battery. USB C Cable is provided for optionally uploading test results to Reporter PC application.

Device Configuration and Setup

Install Battery



- Remove battery door with thumb/finger at arrow
- Insert the CR123A battery, match +/- to indicators



Apply Gray Foam



- Remove brown liner (stickier), align with device.
- Remove white liner and keep for protection of foam when not in use.


Device Configuration and Setup

Test Mode



- Press button to power on device.
- Hold button 5 seconds, view Test Mode screen.
- Blue button shows selected. Change selection if needed

View Device Information



- Select Next from Test Mode screen, view Device information
- Serial number matches number printed on barcode inside battery compartment.

Step 1a: Position Patient



The patient should be in **comfortable** position that allows for relaxation of the leg and foot - it is important that the patient remains **relaxed** during the test.

Make sure you have **access** to the outer ankle bone and the calf.

Make sure you can **see the midline** of the calf.

Patient Positioning Techniques

Lying Methods



Position – side





Position – side



Position - prone

Chair Method



Check Your Position

Tester:


- Patient's outer ankle bone and Achilles tendon are visible
- Stable grip on device

Patient:

- Fully relaxed upper and lower leg


Conditions above not met? Adjust position or try a different one.

Anatomical Landmarks



Preview: Device Alignment

The long probe should align with the outer ankle bone and placed between the ankle bone and the Achilles tendon.



Ensure that device is aligned to the midline.


Ensure that blue arrow is pointing to the back of the knee.

Step 1a: Skin Preparation



Vigorously scrub the test area with the preparation pad provided.

Remove any dirt, lint, moisturizer, etc.



alcohol

Step 1b: Mark Anatomical Landmarks



Midline to Calf

Midline to Calf


Between ankle
bone and the
Achilles tendon

Between ankle
bone and the
Achilles tendon

**Use a skin marker
to draw the marks
shown in the
image.**



Step 2: Power On





- Power on the device by pressing the power button
- The display will prompt you to connect the biosensor

Step 3: Insert Biosensor




- Remove the white liner from gray foam (if not already done) and set aside for later use.
- **Fully** insert the biosensor into the port.
- Align the biosensor to the foam
- The display will prompt you to select the limb (next slide).
- *Tip: Align the biosensor with the foam on all sides, “REMOVE” label side faces up.*

Step 4: Select Limb



- Select the leg to be tested on the touch screen.
- Once selected, “Ready to Test:” will be displayed followed by the limb selected (and green LED will illuminate).
- The back arrow allows you to return to the previous screen if needed.



Step 5: Apply Gel



- Apply a small amount of Signa gel to each probe*.
- The head of the probe should be covered with gel.
- *Tip: Remove excess gel that may lead to gel smearing between the probes.*

**Note: Must use the Signa gel provided with biosensors.*

Step 6: Remove Biosensor Backing



- Remove the backing from the biosensor*.

**Note: Save the foam liner or biosensor liner to protect the foam when device is not in use.*

Step 7: Place Probes



- Place the long probe at the mark made between center of outer ankle bone and Achilles tendon.
- Do not press down yet.
- *Tip: The probes should be behind and not over the ankle bone.*

Step 8: Place Biosensor



- Align the edge of the biosensor to the mark made along the center of the Achilles tendon.
- Arrow closest to Achilles will point toward back of the knee.
- Do not press down yet.

Step 9: Finalize Placement and Contact



- Check your placement
 - Is long probe **behind** Ankle Bone?
 - Is edge of biosensor along **center** of Achilles?
- Check that gel has not smeared between probes
- Press down on **probes** – both probes contact fully
- Press down on **foam** – biosensor contacts fully
- Make sure your **grip is stable** and **pressure is even** - hold with two hands if needed

Step 10: Start Test



- Press the button to start.
- The display will show “Testing”.
- During the test, the LED blinks green when each stimulus is delivered.
- Maintain constant pressure during the test.
- Test time will vary but generally lasts for 10-15 seconds.

Step 11: View Results



- Once results are displayed, the test is complete.
- Remove device from the leg.
- **Amp** shows the amplitude of the sural nerve response in microvolts.
- **CV** shows the conduction velocity of the sural nerve response in meters per second.
- **Limb** tested shown in upper right of screen.


Optional Review: Waveform

- To view the waveform of the measured sural response, press area where Amp and CV results are displayed.
- When you touch that area, it will highlight with an outline.
- Once you remove your finger, the waveform will display.
- Press in the waveform area to return to the Amp/CV results.



Optional Review: Waveform

- For small amplitude waveforms, scaling may be adjusted.



Test Complete


- When results are shown, the test is complete.
- Record the results directly from the device, or upload to the Reporter application.
- To start a new test right away, press the button or insert a new biosensor.
- Device will power off after 1-2 minutes of inactivity.

Testing Protocol

- The test will provide a nerve conduction result the first time in most patients.
- If the first test does not provide a result, the test should be repeated.
 - Pressing the test button again is usually all that is required.
- If the repeat test does not provide a result, the opposite leg should be tested.
 - The same biosensor may be used on both legs.
- There will be a small percentage of patients that you will not be able to obtain results on.



Review Results

- After device powers off, you can still review your most recent results:
 - Power on device with no biosensor connected to show Home screen.
 - Blue outline and checkmark indicate which limb has a result.
 - Press button or touch limb to navigate to results.
 - Press button to return to Home screen.





Low Amplitude Results

- In some cases where amplitude of sural nerve response is low, only Amp will be reported.
- These are valid test results.






Confirming Limb

- In some cases, you may be prompted to confirm the limb during the test.
- If limb matches original selection, results are reported.
- If limb does not match original selection, test must be repeated.




Error Results

- DPNCheck 2.0 has a number of internal checks to ensure that data quality is sufficient for reporting results.
- Error message will be shown if results cannot be reported.
- Help info screen  contains troubleshooting information (also included in Reference Guide card).




Battery


- DPNCheck 2.0 uses the Panasonic CR123A battery (replacements can be ordered through NeuroMetrix).
- Single battery should last at least 100 tests.
- Battery icon shows when to replace.



Battery OK



Battery Low,
Replace soon



Replace Immediately*


**When battery very low, device
may not power on.*

Biosensor Icon

- DPNCheck 2.0 indicates proper biosensor connection



No Biosensor Connected



Biosensor Connected

Cleaning/Re-Use

- The biosensor may be used to test other leg on the same patient if needed.
- The biosensor must be changed between patients.
- The device probes should be cleaned between patients
 - Wipe away excess gel
 - Isopropyl Alcohol wipes may be used to clean/disinfect probes
 - Check temperature lens between probes for gel and wipe away
- The foam does not contact the patient and should not need cleaning unless contaminated. Replace the foam when no longer sticky.
- A soft cloth or wipe with water or isopropyl alcohol may be used to clean the unit exterior as needed; do not use abrasive cleaners or strong solvents.

Documentation & Reporting

Reporting Documentation & Interpretation

- **Report Generation Software:**
 - Generate report (PDF, HL7 or XML)
 - Automatic comparison to normal limits
 - Archive data
- **NCS waveforms and values provide detailed documentation of neuropathy status**
- **Currently relies on manual data entry, development underway for EHR integration solution.**

Sural Nerve Conduction Study


Thomas Kelvin, MD
Technician: Sarah Thomas
Main Street Primary Care
100 Main St
Boston, MA 01234

Patient Name: Robert Fairfield
Patient ID: 3337654
Gender: Male
Birth Date: 3/3/1933


Test Date: 12/9/2020 8:26 PM

	Right	Normal Limit
Test Date	12/9/2020	
Amplitude (μ V)	26	3
Conduction Velocity (m/s)	58	44

Interpretation Guide



Right



Interpretation: ☒ Normal ☐ Mild DPN ☐ Moderate DPN ☐ Severe DPN ☐ Other _____

Nerve conduction study of the right sural nerve is normal.

Thomas Kelvin, MD _____ Date _____

Resources

- DPNCheck 2.0 User Manual (Model NC-040) is available on the website:
www.dpncheck.com/dpncheck-user-manual
- The following materials are provided with your DPNCheck device:
 - Reference Guide with quick-start instructions
 - Patient Positioning Guide
 - Interpretation Guide (without age/height adjustment)

Contact Information

For immediate help when testing contact:

NeuroMetrix Customer Service @ (888) 786-7287

Questions and discussion



APPENDIX: Documentation & Research

DPNCheck® has been clinically validated

- Published data on thousands of patients

Over 30 peer-reviewed clinical studies and review articles



Respiratory Research



DPNCHECK DETECTS DIABETIC PERIPHERAL NEUROPATHY WITH HIGH SENSITIVITY AND SPECIFICITY

Study Publication	Type 2	Type 1	No Diabetes	Total	Reference Diagnosis	Sensitivity	Specificity
Binns-Hall et al. 2018	231	5	0	236	Clinical	0.84	0.68
Papanas et al. 2019	0	53	0	53	Clinical	0.96	0.93
Chatzikosma et al. 2016	114	0	46	160	Clinical	0.91	0.86
Hirayasu et al. 2018	92	0	0	92	Clinical	0.85	0.86
Lee et al. 2014	28	16	0	44	NCS	0.95	0.71
Kural et al. 2018	168	0	0	168	NCS	0.82	0.85
Scarr et al. 2018	0	68	71	139	NCS	0.86	0.79
Total	633	142	117	892		0.88*	0.82*

*Summary sensitivity and specificity determined by bivariate meta-analysis.

9 out of 10 polyneuropathy cases detected